

# Certificate of Analysis

**Product** EXIGO LYSE (1.9 L)  
**Product No** 1504202  
**Lot No** 2308-950  
**Manufacturing Date** 2023-08-01  
**Expiration Date** 2026-07-31

| Test                        | Specification |
|-----------------------------|---------------|
| pH                          | 6.95 – 7.25   |
| Osmolality (mOsm/kg)        | 195 - 225     |
| Conductivity (mS/cm)        | 12.10 – 12.80 |
| CBC Performance *           | Passes test   |
| Particles (PLT/RBC channel) | ≤ 5           |

\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual values meet specifications.

Date: 2023-08-10



*Sofia Karlström*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

