

Certificate of Analysis

Product SWELAB LUMI-L2 LYSE (500 mL)
Product No 1504516
Lot No 2023061901
Manufacturing Date 2023-06-19
Expiration Date 2025-06-18

Test	Specification
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC (x 10 ⁹ /L)	≤ 0,2

*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2023-09-13



Sofia Karlström
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

