

Certificate of Analysis

Product SWELAB ALFALYSE, RFID (5 L)
Product No 1504463
Lot No 2309-979
Manufacturing Date 2023-09-24
Expiration Date 2026-09-23

Test	Specification
pH	7,80 – 8,30
Osmolality (mOsm/kg)	295 – 330
Conductivity (mS/cm)	17,75 – 18,25
CBC Performance *	Passes test
Particles (WBC channel)	≤ 0,2

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.

Date: 2023-09-28



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

