

Certificate of Analysis

Product EXIGO DILUENT
Product No 1504328 (10 L)
Lot No 2309-983
Manufacturing Date 2023-10-02
Expiration Date 2026-10-01

Test	Specification
pH	6.90 – 7.20
Osmolality (mOsm/kg)	300 – 330
Conductivity (mS/cm)	18.00 - 19.00
Particles (PLT/RBC channel)	≤ 5

Actual values meet specifications.

Date: 2023-10-06



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

