


Certificate of Analysis

Product H50V Probe Cleaner
Product No 1440111 (100 mL)
Lot No 23070031M4
Manufacturing Date 2023-07-27
Expiration Date 2024-07-26

Test	Specification
Appearance	Transparent liquid without precipitation, particle or floccules
pH	12,00 - 14,00
Blank	WBC $\leq 0.5 \times 10^9/L$ RBC $\leq 0.1 \times 10^{12}/L$ HGB $\leq 5g/L$ PLT $\leq 20 \times 10^9/L$

Actual values meet specifications.

Date: 2023-10-09



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

