

Certificate of Analysis

Product	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
Product No	1070030 1070039
Lot No	23114540
Manufacturing Date	2023-08-24
Expiration Date	2026-07-28

Parameter	Specification
Volume 20 μ L	CV \leq 1,0%
WBC background*	\leq 0,1 x 10 ⁹ /L
RBC background*	\leq 0,01 x 10 ¹² /L
HGB background	\leq 0,2 g/dL
PLT background	\leq 10 x 10 ⁹ /L
*WBC background for Exigo	\leq 0,2 x 10 ⁹ /L
*RBC background for Exigo	\leq 0,03 x 10 ¹² /L

Actual value meets specification.

Date: 2023-10-12



Ulrika Miefalk
Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

