

# Certificate of Analysis

**Product** SWELAB ALFADILUENT, RFID (20 L)  
**Product No** 1504462  
**Lot No** 2310-985  
**Manufacturing Date** 2023-10-09  
**Expiration Date** 2026-10-08

| Test                        | Specification |
|-----------------------------|---------------|
| pH                          | 6,70 – 6,90   |
| Osmolality (mOsm/kg)        | 294 – 312     |
| Conductivity (mS/cm)        | 16,98 – 17,93 |
| Particles (PLT/RBC channel) | ≤ 5           |

Actual value meets specification.

Date: 2023-10-17



Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

