

Certificate of Analysis

Product H50V SHEATH, RFID
Product No 1440108 (1.9 L)
Lot No 23080635M4
Manufacturing Date 2023-09-27
Expiration Date 2025-09-26

Test	Specification
Appearance	Sheath should be transparent liquid without precipitation, particles or floccules
pH	8.10 - 8.70
Osmolality (mOsm/kg)	20 - 60
Blanks	WBC $\leq 0.2 \times 10^9/L$

Actual values meet specifications.

Date: 2023-10-17



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

