

Certificate of Analysis


Product MEDONIC M-SERIES LYSE, RFID (5 L)
Product No 1504461
Lot No 2310-988
Manufacturing Date 2023-10-16
Expiration Date 2026-10-15

Test	Specification
pH	6,80 – 7,20
Osmolality (mOsm/kg)	220 – 240
Conductivity (mS/cm)	14,00 – 15,00
CBC Performance test*	Passes test
Particles (WBC channel)	≤ 0,2

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.

Date: 2023-10-20



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

