

Certificate of Analysis

Product H50V DILUENT, RFID
Product No 1440103 (5 L)
Lot No 23090631M2
Manufacturing Date 2023-09-26
Expiration Date 2025-09-25

Test	Specification
Appearance	Diluent should be transparent liquid without precipitation, particle or floccules
pH	6.80 – 7.00
Osmolality (mOsm/kg)	309 – 321
Conductivity (mS/cm)	18.50 - 19.50
Particles (10 ⁵ /L) (Particle volume ≥ 2.5fL)	≤ 2.5

Actual values meet specifications.

Date: 2023-10-20



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

