

# Certificate of Analysis


<b>Product</b>	MEDONIC M51-L1 LYSE (200 mL)
<b>Product No</b>	1504511
<b>Lot No</b>	2023053001
<b>Manufacturing Date</b>	2023-05-30
<b>Expiration Date</b>	2025-05-29

<b>Test</b>	<b>Specification</b>
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC (x 10 <sup>9</sup> /L)	≤ 0,2

\*Performance testing at a reference laboratory. WBC and HGB tested and approved.

Actual values meet specification.

Date: 2023-10-18

  
Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

