

Certificate of Analysis

Product	MEDONIC M51-L2 LYSE (500 mL)
Product No	1504512
Lot No	2023061901
Manufacturing Date	2023-06-19
Expiration Date	2025-06-18

Test	Specification
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC ($\times 10^9/L$)	$\leq 0,2$

*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2023-10-18



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

