a Boule Diagnostics, AB company



CORPORATE SEAL 1998 FLORIDA

CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Stopper

Lot No: 2309-201

Product Part #: 501-350

Exp. Date: 2026-08-31

Manufacturing Date: 09-13-2023

Quality Testing Date: 09-15-2023

Parameter	Specification	Result	Pass/Fail
pH (units):	10.10 – 11.20	11.16	Pass
Osmolality (mOsm/kg):	830 – 860	843	Pass
Conductivity (mS/cm):	42.7 – 43.3	43.1	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Quality and Regulatory

Date