

# Certificate of Analysis

**Product** MEDONIC M51-L2 LYSE (500 mL)  
**Product No** 1504512  
**Lot No** 2023071201  
**Manufacturing Date** 2023-07-12  
**Expiration Date** 2025-07-11

Test	Specification
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC (x 10 <sup>9</sup> /L)	≤ 0,2

\*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2023-11-14



*Ulrika Miefalk*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

