


Certificate of Analysis

Product BOULE EASYCLEANER (50 mL)
Product No 1504513
Lot No. 2023102301
Manufacturing date 2023-10-23
Expiration Date 2025-10-22

Test Hypochlorite	Specification
Hypochlorite concentration	4.50 – <5 %
pH	12.60 – 13.20
Blank	PLT $\leq 10 \times 10^9/L$ RBC $\leq 0.02 \times 10^{12}/L$

Actual values meet specifications.

Date: 2023-11-22



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

