

# Certificate of Analysis


**Product** MEDONIC M51-L1 LYSE (200 mL)  
**Product No** 1504511  
**Lot No** 2023081601  
**Manufacturing Date** 2023-08-16  
**Expiration Date** 2025-08-15

Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC (x 10 <sup>9</sup> /L)	≤ 0,2

\*Performance testing at a reference laboratory. WBC and HGB tested and approved.

Actual values meet specification.

Date: 2023-11-30



Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

