


# Certificate of Analysis

**Product** H50V LYSE, RFID  
**Product No** 1440110 (200 mL)  
**Lot No** 23070511M2  
**Manufacturing Date** 2023-08-04  
**Expiration Date** 2025-08-03

Test	Specification
Appearance	Clear liquid, no precipitate particles or flocs
pH	2.00 – 4.50
Absorbance $\lambda_{\max}$ (nm)	530 – 550
Blank	WBC $\leq 0.3 \times 10^9/L$ HGB $\leq 2g/L$
WBC	a) Has two population peaks b) Meet the peak form and position mark requirements.

Actual values meet specifications.

Date: 2023-12-01



*Caroline Sparf*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

