


# Certificate of Analysis

**Product** H50V DETERGENT, RFID  
**Product No** 1440106 (1.9 L)  
**Lot No** 23070377M5  
**Manufacturing Date** 2023-08-04  
**Expiration Date** 2025-08-03

Test	Specification
Appearance	Detergent should be transparent liquid without precipitation, particle or floccules
pH	9.0 – 10.5
Blank	WBC $\leq 0.2 \times 10^9/L$ RBC $\leq 0.02 \times 10^{12}/L$ HGB $\leq 1g/L$ PLT $\leq 10 \times 10^9/L$

Actual values meet specifications.

Date: 2023-12-01



*Caroline Sparf*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

