

---

# Certificate of Analysis

**Product** EXIGO EOS REAGENT (1.9 L)  
**Product No** 1504278  
**Lot No** 2312-119  
**Manufacturing Date** 2023-12-04  
**Expiration Date** 2025-12-03


---

Test	Specification
pH	8.70 – 9.30
Osmolality (mOsm/kg)	125 - 145
Conductivity (mS/cm)	6.80 – 7.60
EOS Performance *	Passes test
Blank EOS	≤ 0.1

\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory.

Actual values meet specifications.

Date: 2023-12-14



*Caroline Sparf*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

