


# Certificate of Analysis

|                           |   |
|---------------------------|---|
| <b>Product</b>            | Boule Micropipette Plastic EDTA 10x100 pcs<br>Boule Micropipette Plastic EDTA 1x100 pcs |
| <b>Product No</b>         | 1070030<br>1070039  |
| <b>Lot No</b>             | 231545A6  |
| <b>Manufacturing Date</b> | 2023-10-11  |
| <b>Expiration Date</b>    | 2026-09-28  |

| <b>Parameter</b>          | <b>Specification</b>              |
|---------------------------|-----------------------------------|
| Volume 20 $\mu$ L         | CV $\leq$ 1,0%                    |
| WBC background*           | $\leq$ 0,1 x 10 <sup>9</sup> /L   |
| RBC background*           | $\leq$ 0,01 x 10 <sup>12</sup> /L |
| HGB background            | $\leq$ 0,2 g/dL                   |
| PLT background            | $\leq$ 10 x 10 <sup>9</sup> /L    |
| *WBC background for Exigo | $\leq$ 0,2 x 10 <sup>9</sup> /L   |
| *RBC background for Exigo | $\leq$ 0,03 x 10 <sup>12</sup> /L |

Actual value meets specification.

Date: 2024-01-16



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

