

Certificate of Analysis


Product SWELAB LUMI-L2 LYSE (500 mL)
Product No 1504516
Lot No 2023102101
Manufacturing Date 2023-10-21
Expiration Date 2025-10-20

Test	Specification
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC (x 10 ⁹ /L)	≤ 0,2

*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2024-01-19



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

