


Certificate of Analysis

Product H50V DETERGENT, RFID
Product No 1440105 (5 L)
Lot No 23100791M4
Manufacturing Date 2023-10-30
Expiration Date 2025-10-29

Test	Specification
Appearance	Detergent should be transparent liquid without precipitation, particle or floccules
pH	9.0 – 10.5
Blank	WBC $\leq 0.2 \times 10^9/L$ RBC $\leq 0.02 \times 10^{12}/L$ HGB $\leq 1g/L$ PLT $\leq 10 \times 10^9/L$

Actual values meet specifications.

Date: 2024-01-25



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

