

Boule Con-5Diff G2

CONTROL

IVD



Control	Lot
Tri-Level	62401-1K
Low	62401-11
Normal	62401-12
High	62401-13

2024-06-05
2024-06-05
2024-06-05
2024-06-05

Open vial stability:

14 days ¹

Quintus

		LOT L 62401-11		LOT N 62401-12		LOT H 62401-13	
		Mean ²	± Range ³	Mean ²	± Range ³	Mean ²	± Range ³
WBC	10 ⁹ /L	4.4	0.5	8.9	0.8	21.4	2.0
NEU#	10 ⁹ /L	2.6	0.5	4.3	1.3	8.3	3.4
LYM#	10 ⁹ /L	1.5	0.4	3.8	0.8	10.9	2.7
MON#	10 ⁹ /L	0.2	0.3	0.4	0.7	1.0	1.9
EO#	10 ⁹ /L	0.1	0.3	0.4	0.3	1.2	0.6
BAS#	10 ⁹ /L	0.0	0.2	0.0	0.2	0.0	0.2
NEU	%	57.8	9.0	47.8	8.0	38.8	8.0
LYM	%	33.7	9.9	43.0	9.9	50.9	9.9
MON	%	5.0	8.7	5.0	8.5	4.7	8.3
EO	%	3.3	4.5	4.0	3.0	5.4	3.6
BAS	%	0.2	1.0	0.2	1.0	0.2	1.0
RBC	10 ¹² /L	2.35	0.15	4.33	0.18	4.88	0.30
Hgb	g/dl	5.7	0.5	12.0	0.6	17.4	0.7
Hct	%	15.9	2.5	33.0	3.0	44.8	3.5
MCV	fL	67.5	5.5	76.1	5.5	91.8	5.5
MCH	pg	24.3	2.8	27.7	2.8	35.7	2.8
MCHC	g/dl	35.8	3.6	36.4	3.0	38.8	3.0
RDWcv	%	19.3	5.0	19.0	5.0	15.7	5.0
Plt	10 ⁹ /L	87	25	230	40	489	65
MPV	fL	10.5	3.0	10.7	3.0	11.0	3.0

i	CONTROL	N
	Target X	Target Y
	91	113

	1
bg	Трайност на отворен флакон
hr	Stabilnost otvorene bočice
cs	Stabilita otevřené ampulky
da	Åben rør stabilitet
et	Stabiilsus avatud viaalis
fr	Stabilité en flacon ouvert
de	Stabilität geöffneter Flaschen
el	Διατήρηση δείγματος μετά την αποσφράγιση
it	Stabilità della fiala aperta
lv	Atvērt flakonu stabilitāti
lt	Stabilumas atidarius buteliuką
no	Åpen rør/glass stabilitet
pl	Trwałość po otwarciu fioleki
pt	Estabilidade após abertura do frasco
ro	Termenul de valabilitate al fiolei desfăcută
ru	Стабильность вскрытого флакона
sr	Proizvod stabilan nakon otvaranja pakovanja
sk	Stabilita otvorenej injekčnej liekovky
sl	Stabilnost odprte stekleničke
es	Estabilidad de la cápsula abierta
sv	Hållbarhet för öppnad flaska
tr	Açık şişe stabilitesi

	2	3
14 дни	Средно	Обхват
14 dani	Sredstvo	Raspon
14 dny	Průměr	Rozsah
14 dage	Middelværdi	Område
14 päevade arv	Keskmine	Vahemik
14 jour	Moyenne	Intervalle
14 tage	Mittelwert	Bereich
14 ημέρα	Μέση	Αναμενόμενο εύρος α
14 giorni	Media	Intervallo
14 dienas	Nozīmē	Diapazons
14 d.	Vidurkis	Intervalas
14 dager	Middelverdi	Område
14 Liczba dni	Średnia	Zakres
14 dias	Média	Intervalo
14 zile	Medie	Interval
14 дней	Среднее	диапазон
14 dana	Srednja vrednost	Opseg
14 dní	Priemer	Rozsah
14 dni	Povprečno	Območje
14 dias	Media	Intervalo
14 dagar	Medelvärde	Intervall
14 gün	Ortalama	Aralığı



www.boule.com

Boule
Boule Medical AB
Domnarvsgatan 4
SE-163 53 Spånga, Sweden

26577-5

1504391

For Quintus with barcode reader

1. In the Main Menu go to "Calibration QC" / "Enter QC reference".
2. Highlight the first Barcode field, scan in each calibrator/control barcode of the column in sequence (1-9) and click "OK".
3. Repeat steps 1 and 2 for each lot to be entered.

Instr.: Quintus
Lot: 62401-11
Prod.: Con-diff Low
ID: 6240111+
Exp.: 5-Jun-2024

Instr.: Quintus
Lot: 62401-12
Prod.: Con-diff Normal
ID: 6240112+
Exp.: 5-Jun-2024

Instr.: Quintus
Lot: 62401-13
Prod.: Con-diff High
ID: 6240113+
Exp.: 5-Jun-2024

