

CERTIFICATE OF ANALYSIS

Product: Swelab™ Alfalyse, 5L

Lot No: 2401-002

Product Part #: 501-207


Exp. Date: 2026-12-31

Manufacturing Date: 2024-01-10

Quality Testing Date: 2024-01-18

Parameter	Specification	Result	Pass/Fail
pH (units):	7.54 – 8.30	8.29	Pass
Osmolality (mOsm/kg):	295 – 330	318	Pass
Conductivity (mS/cm):	18.00– 18.50	18.36	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality Assurance

1/29/24

Date

