

CERTIFICATE OF ANALYSIS

Product: Swelab™AlfaDiluent

Lot No: 2309-001

Product Part #: 501-208

Exp. Date: 2026-08-31

Manufacturing Date: 09-15-2023

Quality Testing Date: 09-19-2023

Parameter	Specification	Result	Pass/Fail
pH (units):	6.50 – 6.90	6.80	Pass
Osmolality (mOsm/kg):	294 – 312	300	Pass
Conductivity (mS/cm):	16.98 – 17.93	17.25	Pass
Background (x10 ³ /μL):	0.0 – 8.0	0.5	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality Assurance

09/25/23

Date

