

CERTIFICATE OF ANALYSIS

Product: Boule Cleaning Kit

Kit Lot No: 1952

Product Part #: 501-036

Kit Exp. Date: 2026-01-31

Part Number: 503-085			
Boule Detergent Cleaner Lot # 1949			
Manufacturing Date: 2024-01-10		Quality Testing Date: 2024-01-16	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.28	Pass


Part Number: 503-086			
Boule Hypochlorite (2%) Cleaner Lot # 1950			
Manufacturing Date: 2024-01-09		Quality Testing Date: 2024-01-10	
Parameter	Specification	Result	Pass/Fail
pH (units):	11.80 – 13.00	12.95	Pass

Part Number: 503-087			
Boule Enzymatic Cleaner Lot # 1951			
Manufacturing Date: 2024-01-08		Quality Testing Date: 2024-01-12	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.08	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality and Regulatory



Date

