

# Certificate of Analysis

**Product** MEDONIC M-SERIES LYSE (5 L)  
**Product No** 1504123  
**Lot No** 2402-150  
**Manufacturing Date** 2024-02-19  
**Expiration Date** 2027-02-18

---

Test	Specification
pH	6,80 – 7,20
Osmolality (mOsm/kg)	220 – 240
Conductivity (mS/cm)	14,00 – 15,00
CBC Performance test*	Passes test
Particles (WBC channel)	≤ 0,2

\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.

Date: 2024-02-26



*Ulrika Miefalk*  
Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

