

CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Lyse

Lot No: 2401-301

Product Part #: 501-291

Exp. Date: 2025-07-31

Manufacturing Date: 2024-01-19

Quality Testing Date: 2024-02-27

Parameter	Specification	Result	Pass/Fail
pH (units):	2.7 – 3.0	3.0	Pass
Osmolality (mOsm/kg):	35 – 44	37	Pass
Conductivity (µS/cm):	1,117 – 1,541	1,216	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Regulatory and Compliance



Date

