

LOT: 12402-14
EXP: 2024-05-07
Open Vial Stability 5 days

EXIGO

	WBC 10 ⁹ /L	RBC 10 ¹² /L	Hgb g/dL	MCV fL	PLT 10 ⁹ /L
System specific values	8.5	4.17	10.7	71.2	215
Tolerance limits ±	0.3	0.10	0.2	2.0	15

This assay sheet is also available on :
www.boule.com

For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo

Lot: 12402-14

Prod.: Vet-Cal

ID: 1240214+

Exp.: 7-May-2024

1:



2:



3:



4:



5:



6:



7:



8:



9:

