

Boule Cal

CAL

IVD



Calibrator

Lot	
22402-34	2024-05-19

Open vial stability 5 days ¹

Medonic

Parameter ²	RBC 10 ¹² /L	MCV fL	Plt 10 ⁹ /L	MPV fL	WBC 10 ⁹ /L	Hgb g/dL
Medonic M-series / M32	4.19	81.9	234	11.6	8.5	12.3
Range ± ³	0.10	2.0	15	1.5	0.3	0.2

Swelab

Parameter ²	RBC 10 ¹² /L	MCV fL	Plt 10 ⁹ /L	MPV fL	WBC 10 ⁹ /L	Hgb g/dL
Swelab Alfa / Alfa Plus	4.19	83.3	239	12.5	8.5	12.3
Range ± ³	0.10	2.0	15	1.5	0.3	0.2

	1	2	3
bg	Трайност на отворен флакон	5 дни	Обхват
hr	Stabilnost otvorene bočice	5 dani	Raspon
cs	Stabilita otevřené ampulky	5 dny	Rozsah
da	Åben rør stabilitet	5 dage	Område
et	Stabiilsus avatud viaalis	5 päevade arv	Vahemik
fr	Stabilité en flacon ouvert	5 jour	Intervalle
de	Stabilität geöffneter Flaschen	5 tage	Bereich
el	Διατήρηση δείγματος μετά την αποσφράγιση	5 ημέρα	Αναμενόμενο εύρος α
it	Stabilità della fiala aperta	5 giorni	Intervallo
lv	Atvērt flakonu stabilitāti	5 dienas	Diapazons
lt	Stabilumas atidarius buteliuką	5 d.	Intervalas
no	Åpen rør/glass stabilitet	5 dager	Område
pl	Trwałość po otworzeniu fiolki	5 Liczba dni	Zakres
pt	Estabilidade após abertura do frasco	5 dias	Intervalo
ro	Termenul de valabilitate al fiolei desfăcute	5 zile	Interval
ru	Стабильность вскрытого флакона	5 дней	диапазон
sr	Proizvod stabilan nakon otvaranja pakovanja	5 dana	Opseg
sk	Stabilita otvorenej injekčnej liekovky	5 dní	Rozsah
sl	Stabilnost odprte stekleničke	5 dni	Območje
es	Estabilidad de la cápsula abierta	5 dias	Intervalo
sv	Hållbarhet för öppnad flaska	5 dagar	Intervall
tr	Açık şişe stabilitesi	5 gün	Aralığı



www.boule.com

Boule
Boule Medical AB
Domnarvsgatan 4
SE-163 53 Spånga, Sweden

25123-6

1504055

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series

Lot: 22402-34

Prod.: Cal

ID: 2240234+

Exp.: 19-May-2024



For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 22402-34
Prod.: Cal
ID: 2240234+
Exp.: 19-May-2024

