

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	--	--
Low	22402-41	2024-07-04
Normal	22402-42	2024-07-04
High	22402-43	2024-07-09

Open vial stability 14 days ¹

MEDONIC

Medonic M-series/M32		CONTROL	
		LOT	22402-41
		Mean ²	± Range ³
RBC	10 ¹² /L	2.24	0.12
MCV	fL	67.2	5.0
RDW	%	14.4	4.0
Hct	%	15.1	2.5
Plt	10 ⁹ /L	77	15
MPV	fL	11.6	1.5
WBC	10 ⁹ /L	3.5	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	25.0	1.5
MCHC	g/dL	37.1	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	45.5	7.0
Mid	%	5.0	5.0
Gran	%	49.5	7.0

CONTROL	LOT	N	
		22402-42	
		Mean ²	± Range ³
		4.21	0.18
		80.8	5.0
		14.5	4.0
		34.0	3.0
		234	30
		11.7	1.5
		8.6	0.6
		12.1	0.4
		28.7	2.5
		35.6	3.0
		4.0	0.6
		0.5	0.5
		4.1	1.0
		46.5	5.0
		5.3	5.0
		48.2	5.0

CONTROL	LOT	H	
		22402-43	
		Mean ²	± Range ³
		5.09	0.22
		91.0	5.0
		14.1	4.0
		46.3	3.5
		524	60
		12.0	1.5
		20.8	1.8
		16.0	0.5
		31.4	1.5
		34.6	3.0
		12.8	1.5
		1.0	1.0
		7.0	2.5
		61.4	5.0
		5.0	5.0
		33.6	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dani (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud vialis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διάρκηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčneje liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

	Средно
	Sredstvo
	Průměr
	Middelverdi
	Keskmine
	Moyenne
	Mittelwert
	Μέση
	Media
	Nozīmē
	Vidurkis
	Middelverdi
	Średnia
	Média
	Medie
	Среднее
	Srednja vrednost
	Priemer
	Povprečno
	Media
	Medelvärde
	Ortalama

3

	Обхват
	Raspon
	Rozsah
	Område
	Vahemik
	Intervalle
	Bereich
	Αναμενόμενο εύρος α
	Intervallo
	Diapazons
	Intervalas
	Område
	Zakres
	Intervalo
	Interval
	диапазон
	Opseg
	Rozsah
	Območje
	Intervalo
	Intervall
	Aralığı

For Medonic M-series / M32 with barcode reader















M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.:	M-series	Instr.:	M-series	Instr.:	M-series
Lot:	22402-41	Lot:	22402-42	Lot:	22402-43
Prod.:	Con-Diff Low	Prod.:	Con-Diff Norm	Prod.:	Con-Diff High
ID:	2240241+	ID:	2240242+	ID:	2240243+
Exp.:	4-Jul-2024	Exp.:	4-Jul-2024	Exp.:	9-Jul-2024

1:	 B o u l e C B A V 2	 B o u l e C B A V 2	 B o u l e C B A V 2
2:	 2240241+	 2240242+	 2240243+
3:	 24070400001153030000	 24070400002153030000	 24070900003153030000
4:	 10004030900000000000	 10004031200000000000	 10004030900000000000
5:	 22412672500560315125	 42118808501210434030	 50922910501600546335
6:	 07715116152501537130	 23430117152871535630	 52460120153141534630
7:	 03503016040020301708	 08606040060050504110	 20818128150101007025
8:	 14440455700505049570	 14540465500535048250	 14140614500505033650
9:	 11115140168069960357	 11111372193170522467	 11119838160799413901

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	--	--
Low	22402-41	2024-07-04
Normal	22402-42	2024-07-04
High	22402-43	2024-07-09

Open vial stability 14 days ¹

SWELAB

Swelab Alfa/Alfa Plus	CONTROL		L
	LOT	22402-41	22402-41
	Mean ²	± Range ³	
RBC	10 ¹² /L	2.26	0.12
MCV	fL	69.0	5.0
RDW	%	14.0	4.0
Hct	%	15.6	2.5
Plt	10 ⁹ /L	81	15
MPV	fL	12.7	1.5
WBC	10 ⁹ /L	3.5	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	24.8	1.5
MCHC	g/dL	35.9	3.0
Lymph#	10 ⁹ /L	1.7	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	47.8	7.0
Mid	%	3.3	5.0
Gran	%	48.9	7.0

CONTROL	N	
	LOT	22402-42
	Mean ²	± Range ³
	4.22	0.18
	82.3	5.0
	13.5	4.0
	34.7	3.0
	248	30
	12.7	1.5
	8.6	0.6
	12.1	0.4
	28.7	1.5
	34.9	3.0
	4.3	0.6
	0.3	0.5
	4.0	1.0
	50.0	5.0
	3.6	5.0
	46.4	5.0

CONTROL	H	
	LOT	22402-43
	Mean ²	± Range ³
	5.10	0.22
	92.8	5.0
	13.0	4.0
	47.3	3.5
	544	60
	13.2	1.5
	20.9	1.8
	16.0	0.5
	31.4	1.5
	33.8	3.0
	13.5	1.5
	0.6	1.0
	6.8	2.5
	64.5	5.0
	3.1	5.0
	32.4	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dani (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud vialis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčnej liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

	Средно
	Sredstvo
	Průměr
	Middelverdi
	Keskmine
	Moyenne
	Mittelwert
	Μέση
	Media
	Nozīmē
	Vidurkis
	Middelverdi
	Średnia
	Média
	Medie
	Среднее
	Srednja vrednost
	Priemer
	Povprečno
	Media
	Medelvärde
	Ortalama

3

	Обхват
	Raspon
	Rozsah
	Område
	Vahemik
	Intervalle
	Bereich
	Αναμενόμενο εύρος α
	Intervallo
	Diapazons
	Intervalas
	Område
	Zakres
	Intervalo
	Interval
	диапазон
	Orseg
	Rozsah
	Območje
	Intervalo
	Intervall
	Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 22402-41
Prod.: Con-Diff Low
ID: 2240241+
Exp.: 4-Jul-2024

Instr.: Swelab Alfa
Lot: 22402-42
Prod.: Con-Diff Norm
ID: 2240242+
Exp.: 4-Jul-2024

Instr.: Swelab Alfa
Lot: 22402-43
Prod.: Con-Diff High
ID: 2240243+
Exp.: 9-Jul-2024

