

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	--	--
Low	22403-41	2024-08-02
Normal	22403-42	2024-08-02
High	22403-43	2024-08-06

MEDONIC

Medonic M-series/M32		CONTROL	
		LOT	L
		22403-41	22403-41
		Mean ²	± Range ³
RBC	10 ¹² /L	2.26	0.12
MCV	fL	68.1	5.0
RDW	%	14.6	4.0
Hct	%	15.4	2.5
Plt	10 ⁹ /L	78	15
MPV	fL	11.6	1.5
WBC	10 ⁹ /L	3.3	0.3
Hgb	g/dL	5.7	0.3
MCH	pg	25.2	1.5
MCHC	g/dL	37.0	3.0
Lymph#	10 ⁹ /L	1.5	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.6	0.8
Lymph	%	45.9	7.0
Mid	%	5.0	5.0
Gran	%	49.1	7.0

	CONTROL	
	LOT	N
	22403-42	22403-42
	Mean ²	± Range ³
	4.17	0.18
	84.5	5.0
	13.8	4.0
	35.2	3.0
	232	30
	12.0	1.5
	8.4	0.6
	12.4	0.4
	29.7	2.5
	35.2	3.0
	3.9	0.6
	0.5	0.5
	4.0	1.0
	47.0	5.0
	5.4	5.0
	47.6	5.0

	CONTROL	
	LOT	H
	22403-43	22403-43
	Mean ²	± Range ³
	5.09	0.22
	93.5	5.0
	13.6	4.0
	47.6	3.5
	512	60
	12.0	1.5
	20.8	1.8
	16.2	0.5
	31.8	1.5
	34.0	3.0
	12.8	1.5
	1.0	1.0
	7.0	2.5
	61.5	5.0
	5.0	5.0
	33.5	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dani (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud vialis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčnej liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

	Средно
	Sredstvo
	Průměr
	Middelværdi
	Keskmine
	Moyenne
	Mittelwert
	Μέση
	Media
	Nozīmē
	Vidurkis
	Middelverdi
	Średnia
	Média
	Medie
	Среднее
	Srednja vrednost
	Priemer
	Povprečno
	Media
	Medelvärde
	Ortalama

3

	Обхват
	Raspon
	Rozsah
	Område
	Vahemik
	Intervalle
	Bereich
	Αναμενόμενο εύρος α
	Intervallo
	Diapazons
	Intervalas
	Område
	Zakres
	Intervalo
	Interval
	диапазон
	Opseg
	Rozsah
	Območje
	Intervall
	Intervall
	Aralığı

For Medonic M-series / M32 with barcode reader























M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.:	M-series	Instr.:	M-series	Instr.:	M-series
Lot:	22403-41	Lot:	22403-42	Lot:	22403-43
Prod.:	Con-Diff Low	Prod.:	Con-Diff Norm	Prod.:	Con-Diff High
ID:	2240341+	ID:	2240342+	ID:	2240343+
Exp.:	2-Aug-2024	Exp.:	2-Aug-2024	Exp.:	6-Aug-2024

1:	 Bou l e C B A V 2	 Bou l e C B A V 2	 Bou l e C B A V 2
2:	 2240341+	 2240342+	 2240343+
3:	 24080200001153030000	 24080200002153030000	 24080600003153030000
4:	 10004031100000000000	 10004031200000000000	 10004031200000000000
5:	 22612681500570315425	 41718845501240435230	 50922935501620547635
6:	 07815116152521537030	 23230120152971535230	 51260120153181534030
7:	 03303015040020301608	 08406039060050504010	 20818128150101007025
8:	 14640459700505049170	 13840470500545047650	 13640615500505033550
9:	 11114735190916340197	 11110196174537384907	 11113070690866874331

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	--	--
Low	22403-41	2024-08-02
Normal	22403-42	2024-08-02
High	22403-43	2024-08-06

Open vial stability 14 days ¹

SWELAB

Swelab Alfa/Alfa Plus	CONTROL		L
	LOT	22403-41	22403-41
	Mean ²	± Range ³	
RBC	10 ¹² /L	2.29	0.12
MCV	fL	69.5	5.0
RDW	%	14.3	4.0
Hct	%	15.9	2.5
Plt	10 ⁹ /L	78	15
MPV	fL	12.7	1.5
WBC	10 ⁹ /L	3.3	0.3
Hgb	g/dL	5.8	0.3
MCH	pg	25.3	1.5
MCHC	g/dL	36.5	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.6	0.8
Lymph	%	48.3	7.0
Mid	%	3.3	5.0
Gran	%	48.4	7.0

CONTROL	N	
	LOT	22403-42
	Mean ²	± Range ³
	4.19	0.18
	85.5	5.0
	12.7	4.0
	35.8	3.0
	231	30
	13.0	1.5
	8.4	0.6
	12.5	0.4
	29.8	1.5
	34.9	3.0
	4.2	0.6
	0.3	0.5
	3.9	1.0
	49.9	5.0
	3.3	5.0
	46.8	5.0

CONTROL	H	
	LOT	22403-43
	Mean ²	± Range ³
	5.12	0.22
	94.7	5.0
	12.6	4.0
	48.5	3.5
	525	60
	13.1	1.5
	20.7	1.8
	16.4	0.5
	32.0	1.5
	33.8	3.0
	13.4	1.5
	0.6	1.0
	6.7	2.5
	64.6	5.0
	2.8	5.0
	32.6	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dana (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud vialis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčnej liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

	Средно
	Sredstvo
	Průměr
	Middelverdi
	Keskmine
	Moyenne
	Mittelwert
	Μέση
	Media
	Nozīmē
	Vidurkis
	Middelverdi
	Średnia
	Média
	Media
	Среднее
	Srednja vrednost
	Priemer
	Povprečno
	Media
	Medelvärde
	Ortalama

3

	Обхват
	Raspon
	Rozsah
	Område
	Vahemik
	Intervalle
	Bereich
	Αναμενόμενο εύρος α
	Intervallo
	Diapazons
	Intervalas
	Område
	Zakres
	Intervalo
	Interval
	диапазон
	Orseg
	Rozsah
	Območje
	Intervalo
	Intervall
	Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 22403-41
Prod.: Con-Diff Low
ID: 2240341+
Exp.: 2-Aug-2024

Instr.: Swelab Alfa
Lot: 22403-42
Prod.: Con-Diff Norm
ID: 2240342+
Exp.: 2-Aug-2024

Instr.: Swelab Alfa
Lot: 22403-43
Prod.: Con-Diff High
ID: 2240343+
Exp.: 6-Aug-2024

