

# Certificate of Analysis

**Product** MEDONIC M51-D DILUENT (20L)  
**Product No** 1504510  
**Lot No** 2024032602  
**Manufacturing Date** 2024-03-26  
**Expiration Date** 2026-03-25

Test	Specification
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel (x 10 <sup>9</sup> /L)	≤ 5

Actual values meet specification.

Date: 2024-05-13



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

