

CERTIFICATE OF ANALYSIS

Product: Medonic™ M-Series Lyse

Lot No: 2405-001

Product Part #: 501-212

Exp. Date: 2027-04-30

Manufacturing Date: 2024-05-07

Quality Testing Date: 2024-05-13

Parameter	Specification	Result	Pass/Fail
pH (units):	6.80 – 7.20	7.12	Pass
Osmolality (mOsm/kg):	220 – 240	230	Pass
Conductivity (mS/cm):	14.00 – 15.00	14.09	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Regulatory and Compliance



Date

