

Certificate of Analysis

Product Boule Micropipette Plastic EDTA 10x100 pcs
Boule Micropipette Plastic EDTA 1x100 pcs

Product No 1070030
1070039

Lot No 240145H0


Manufacturing Date 2024-03-18

Expiration Date 2027-02-28

Parameter	Specification
Volume 20 μ L	CV \leq 1,0%
WBC background*	\leq 0,1 x 10 ⁹ /L
RBC background*	\leq 0,01 x 10 ¹² /L
HGB background	\leq 0,2 g/dL
PLT background	\leq 10 x 10 ⁹ /L
*WBC background for Exigo	\leq 0,2 x 10 ⁹ /L
*RBC background for Exigo	\leq 0,03 x 10 ¹² /L

Actual value meets specification.

Date: 2024-06-03



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

