

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	--	--
Low	22405-41	2024-10-04
Normal	22405-42	2024-10-04
High	22405-43	2024-10-08

MEDONIC

CONTROL		L	
LOT		22405-41	
Medonic M-series/M32	Mean ²	± Range ³	
RBC	10 ¹² /L	2.30	0.12
MCV	fL	67.3	5.0
RDW	%	15.1	4.0
Hct	%	15.5	2.5
Plt	10 ⁹ /L	85	15
MPV	fL	11.4	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.7	0.3
MCH	pg	24.8	1.5
MCHC	g/dL	36.8	3.0
Lymph#	10 ⁹ /L	1.5	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	44.3	7.0
Mid	%	5.9	5.0
Gran	%	49.8	7.0

CONTROL		N	
LOT		22405-42	
	Mean ²	± Range ³	
	4.20	0.18	
	86.0	5.0	
	13.8	4.0	
	36.1	3.0	
	236	30	
	11.6	1.5	
	8.7	0.6	
	12.6	0.4	
	30.0	1.5	
	34.9	3.0	
	3.9	0.6	
	0.6	0.5	
	4.2	1.0	
	44.8	5.0	
	7.3	5.0	
	47.9	5.0	

CONTROL		H	
LOT		22405-43	
	Mean ²	± Range ³	
	5.15	0.22	
	94.8	5.0	
	14.2	4.0	
	48.8	3.5	
	525	60	
	12.0	1.5	
	20.9	1.8	
	16.5	0.5	
	32.0	1.5	
	33.8	3.0	
	12.6	1.5	
	1.2	1.0	
	7.1	2.5	
	60.3	5.0	
	5.6	5.0	
	34.1	5.0	

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dni (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αεροστεγήσση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonū stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Sредно
 - Sredstvo
 - Průměr
 - Middelværdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.:	M-series	Instr.:	M-series	Instr.:	M-series
Lot:	22405-41	Lot:	22405-42	Lot:	22405-43
Prod.:	Con-Diff Low	Prod.:	Con-Diff Norm	Prod.:	Con-Diff High
ID:	2240541+	ID:	2240542+	ID:	2240543+
Exp.:	4-Oct-2024	Exp.:	4-Oct-2024	Exp.:	8-Oct-2024

1:	 Bou l e C B A V 2	 Bou l e C B A V 2	 Bou l e C B A V 2
2:	 2240541+	 2240542+	 2240543+
3:	 24100400001153030000	 24100400002153030000	 24100800003153030000
4:	 10004030900000000000	 10004030400000000000	 10004030800000000000
5:	 23012673500570315525	 42018860501260436130	 51522948501650548835
6:	 08515114152481536830	 23630116153001534930	 52560120153201533830
7:	 03403015040020301708	 08706039060060504210	 20918126150121007125
8:	 15140443700595049870	 13840448500735047950	 14240603500565034150
9:	 11113525470193883422	 11118861078277165301	 11119173509155693572

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	--	--
Low	22405-41	2024-10-04
Normal	22405-42	2024-10-04
High	22405-43	2024-10-08

SWELAB

Swelab Alfa/Alfa Plus		CONTROL LOT	L 22405-41
		Mean	± Range
RBC	10 ¹² /L	2.31	0.12
MCV	fL	68.5	5.0
RDW	%	14.4	4.0
Hct	%	15.8	2.5
Plt	10 ⁹ /L	86	15
MPV	fL	12.5	1.5
WBC	10 ⁹ /L	3.3	0.3
Hgb	g/dL	5.7	0.3
MCH	pg	24.7	1.5
MCHC	g/dL	36.1	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.6	0.8
Lymph	%	47.7	7.0
Mid	%	4.2	5.0
Gran	%	48.1	7.0

CONTROL LOT	N 22405-42
Mean	± Range
4.22	0.18
87.5	5.0
12.5	4.0
36.9	3.0
237	30
12.6	1.5
8.6	0.6
12.6	0.4
29.9	1.5
34.1	3.0
4.2	0.6
0.4	0.5
4.0	1.0
48.6	5.0
4.6	5.0
46.8	5.0

CONTROL LOT	H 22405-43
Mean	± Range
5.13	0.22
96.3	5.0
13.3	4.0
49.4	3.5
538	60
13.8	1.5
20.5	1.8
16.3	0.5
31.8	1.5
33.0	3.0
13.1	1.5
0.7	1.0
6.7	2.5
63.8	5.0
3.4	5.0
32.8	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dni (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αλοσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonū stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelværdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 22405-41
Prod.: Con-Diff Low
ID: 2240541+
Exp.: 4-Oct-2024

Instr.: Swelab Alfa
Lot: 22405-42
Prod.: Con-Diff Norm
ID: 2240542+
Exp.: 4-Oct-2024

Instr.: Swelab Alfa
Lot: 22405-43
Prod.: Con-Diff High
ID: 2240543+
Exp.: 8-Oct-2024

