

Certificate of Analysis

Product MEDONIC M-SERIES DILUENT, RFID (20 L)
Product No 1504460
Lot No 2406-208
Manufacturing Date 2024-06-19
Expiration Date 2027-06-18

Test	Specification
pH	6,90 – 7,10
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,50 – 18,00
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2024-06-25



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

