

# Certificate of Analysis

<b>Product</b>	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
<b>Product No</b>	1070030 1070039
<b>Lot No</b>	240145H1
<b>Manufacturing Date</b>	2024-03-19
<b>Expiration Date</b>	2027-02-28

<b>Parameter</b>	<b>Specification</b>
Volume 20 µL	CV ≤ 1,0%
WBC background*	≤ 0,1 x 10 <sup>9</sup> /L
RBC background*	≤ 0,01 x 10 <sup>12</sup> /L
HGB background	≤ 0,2 g/dL
PLT background	≤ 10 x 10 <sup>9</sup> /L
*WBC background for Exigo	≤ 0,2 x 10 <sup>9</sup> /L
*RBC background for Exigo	≤ 0,03 x 10 <sup>12</sup> /L

Actual value meets specification.

Date: 2024-06-05



Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

