

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹



Control	Lot	
Tri-Level	32406-5K	2024-11-02
Low	32406-51	2024-11-02
Normal	32406-52	2024-11-05
High	32406-53	2024-11-08

MEDONIC

		CONTROL	L
		LOT	32406-51
Medonic M-series/M32		Mean ²	± Range ³
RBC	10 ¹² /L	2.22	0.12
MCV	fL	67.3	7.0
RDW	%	15.8	4.5
Hct	%	14.9	3.0
Plt	10 ⁹ /L	81	15
MPV	fL	11.1	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.5	0.3
MCH	pg	24.8	2.0
MCHC	g/dL	36.9	4.0
Lymph#	10 ⁹ /L	1.5	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	44.1	7.0
Mid	%	6.0	5.0
Gran	%	49.9	7.0

		CONTROL	N
		LOT	32406-52
		Mean ²	± Range ³
		4.21	0.18
		78.7	7.0
		15.4	4.5
		33.1	3.5
		228	30
		11.7	1.5
		8.7	0.6
		11.6	0.4
		27.6	2.0
		35.0	4.0
		3.9	0.6
		0.6	0.5
		4.2	1.0
		44.5	5.0
		7.3	5.0
		48.2	5.0

		CONTROL	H
		LOT	32406-53
		Mean ²	± Range ³
		5.08	0.22
		93.2	7.0
		13.7	4.5
		47.3	4.0
		518	60
		11.6	1.5
		20.8	1.8
		16.3	0.5
		32.1	2.0
		34.5	4.0
		12.5	1.5
		1.2	1.0
		7.1	2.5
		60.0	5.0
		5.7	5.0
		34.3	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dny (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Σταθιρότητα δειγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dias (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Sredno
 - Sredstvo
 - Průměr
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Ανεπιβόμητο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı



[http://www.medonic.se/
support/downloads](http://www.medonic.se/support/downloads)



25122-4

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series	Instr.: M-series	Instr.: M-series
Lot: 32406-51	Lot: 32406-52	Lot: 32406-53
Prod.: Con-Diff Low	Prod.: Con-Diff Norm	Prod.: Con-Diff High
ID: 3240651+	ID: 3240652+	ID: 3240653+
Exp.: 2-Nov-2024	Exp.: 5-Nov-2024	Exp.: 8-Nov-2024

1:  Boul e CBAV2	 Boul e CBAV2	 Boul e CBAV2
2:  3240651+	 3240652+	 3240653+
3:  24110200001153030000	 24110500002153030000	 24110800003153030000
4:  10004030800000000000	 10004030100000000000	 10004030800000000000
5:  22212673700550314930	 42118787701160433135	 50822932701630547340
6:  08115111152482036940	 22830117152762035040	 51860116153212034540
7:  03403015040020301708	 08706039060060504210	 20818125150121007125
8:  15845441700605049970	 15445445500735048250	 13745600500575034350
9:  11117864830372900898	 11115250915041860444	 11114290979330790113

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹



Control	Lot	
Tri-Level	32406-5K	2024-11-02
Low	32406-51	2024-11-02
Normal	32406-52	2024-11-05
High	32406-53	2024-11-08

SWELAB

		CONTROL	L
		LOT	32406-51
		Mean	± Range
Swelab Alfa/Alfa Plus			
RBC	10 ¹² /L	2.23	0.12
MCV	fL	69.0	7.0
RDW	%	15.1	4.5
Hct	%	15.4	3.0
Plt	10 ⁹ /L	80	15
MPV	fL	12.3	1.5
WBC	10 ⁹ /L	3.3	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	25.1	2.0
MCHC	g/dL	36.4	4.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.6	0.8
Lymph	%	47.6	7.0
Mid	%	4.3	5.0
Gran	%	48.1	7.0

		CONTROL	N
		LOT	32406-52
		Mean	± Range
		4.23	0.18
		79.8	7.0
		14.4	4.5
		33.8	3.5
		232	30
		12.8	1.5
		8.7	0.6
		11.7	0.4
		27.7	2.0
		34.6	4.0
		4.2	0.6
		0.4	0.5
		4.1	1.0
		48.7	5.0
		4.7	5.0
		46.6	5.0

		CONTROL	H
		LOT	32406-53
		Mean	± Range
		5.08	0.22
		94.5	7.0
		12.4	4.5
		48.0	4.0
		531	60
		12.6	1.5
		20.6	1.8
		16.2	0.5
		31.9	2.0
		33.8	4.0
		13.1	1.5
		0.7	1.0
		6.8	2.5
		63.5	5.0
		3.6	5.0
		32.9	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Σταθιρότητα δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonū stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcută 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Sredno
 - Sredstvo
 - Průměr
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Intervall
 - диапазон
 - Orseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı



[www.swelab.com/
support/downloads](http://www.swelab.com/support/downloads)

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
 Lot: 32406-51
 Prod.: Con-Diff Low
 ID: 3240651+
 Exp.: 2-Nov-2024

Instr.: Swelab Alfa
 Lot: 32406-52
 Prod.: Con-Diff Norm
 ID: 3240652+
 Exp.: 5-Nov-2024

Instr.: Swelab Alfa
 Lot: 32406-53
 Prod.: Con-Diff High
 ID: 3240653+
 Exp.: 8-Nov-2024

