a Boule Diagnostics, AB company



## **CERTIFICATE OF ANALYSIS**

Product: Quintus 5-Part Lyse

Lot No: 2402-301

Product Part #: 501-291

Exp. Date: 2025-08-31

Manufacturing Date: 2024-02-01

Quality Testing Date: 2024-03-05

Parameter	Specification	Result	Pass/Fail
pH (units):	2.7 – 3.0	3.0	Pass
Osmolality (mOsm/kg):	35 – 44	38	Pass
Conductivity (µS/cm):	1,117 – 1,541	1,228	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Regulatory and Compliance

Date