

# Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days <sup>1</sup>

Control	Lot	
Tri-Level	--	--
Low	22406-41	2024-11-01
Normal	22406-42	2024-11-01
High	22406-43	2024-11-05

## MEDONIC

Medonic M-series/M32		CONTROL	L
		LOT	22406-41
		Mean <sup>2</sup>	± Range <sup>3</sup>
RBC	10 <sup>12</sup> /L	2.21	0.12
MCV	fL	64.2	5.0
RDW	%	14.8	4.0
Hct	%	14.2	2.5
Plt	10 <sup>9</sup> /L	78	15
MPV	fL	11.1	1.5
WBC	10 <sup>9</sup> /L	3.4	0.3
Hgb	g/dL	5.3	0.3
MCH	pg	24.0	1.5
MCHC	g/dL	37.3	3.0
Lymph#	10 <sup>9</sup> /L	1.5	0.4
Mid#	10 <sup>9</sup> /L	0.2	0.3
Gran#	10 <sup>9</sup> /L	1.7	0.8
Lymph	%	44.5	7.0
Mid	%	5.4	5.0
Gran	%	50.1	7.0

CONTROL	N	LOT	22406-42	Mean <sup>2</sup>	± Range <sup>3</sup>
				4.15	0.18
				80.6	5.0
				14.2	4.0
				33.4	3.0
				233	30
				11.6	1.5
				8.6	0.6
				11.9	0.4
				28.7	1.5
				35.6	3.0
				3.9	0.6
				0.6	0.5
				4.1	1.0
				45.0	5.0
				7.4	5.0
				47.6	5.0

CONTROL	H	LOT	22406-43	Mean <sup>2</sup>	± Range <sup>3</sup>
				5.04	0.22
				92.4	5.0
				14.7	4.0
				46.6	3.5
				530	60
				11.7	1.5
				21.3	1.8
				16.2	0.5
				32.1	1.5
				34.8	3.0
				12.9	1.5
				1.2	1.0
				7.2	2.5
				60.5	5.0
				5.6	5.0
				33.9	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dani (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud vialis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčnej liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

	Средно
	Sredstvo
	Průměr
	Middelværdi
	Keskmine
	Moyenne
	Mittelwert
	Μέση
	Media
	Nozīmē
	Vidurkis
	Middelverdi
	Średnia
	Média
	Medie
	Среднее
	Srednja vrednost
	Priemer
	Povprečno
	Media
	Medelvärde
	Ortalama

3

	Обхват
	Raspon
	Rozsah
	Område
	Vahemik
	Intervalle
	Bereich
	Αναμενόμενο εύρος α
	Intervallo
	Diapazons
	Intervalas
	Område
	Zakres
	Intervalo
	Interval
	диапазон
	Opseg
	Rozsah
	Območje
	Intervall
	Aralıqı

# For Medonic M-series / M32 with barcode reader

## M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

## M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series  
Lot: 22406-41  
Prod.: Con-Diff Low  
ID: 2240641+  
Exp.: 1-Nov-2024

Instr.: M-series  
Lot: 22406-42  
Prod.: Con-Diff Norm  
ID: 2240642+  
Exp.: 1-Nov-2024

Instr.: M-series  
Lot: 22406-43  
Prod.: Con-Diff High  
ID: 2240643+  
Exp.: 5-Nov-2024

