

Certificate of Analysis

Product SWELAB ALFADILUENT, RFID (20 L)
Product No 1504462
Lot No 2407-225
Manufacturing Date 2024-07-30
Expiration Date 2027-07-29

Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2024-08-02



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

