

Certificate of Analysis

Product SWELAB ALFA COMBOPACK (1 L Lyse / 5 L Diluent)
Product No 1504127
Lot No 2408-227
Manufacturing Date 2024-08-01
Expiration Date 2027-07-31

Test	Specification Diluent	Specification Lyse
pH	6,70 – 6,90	7,80 – 8,30
Osmolality (mOsm/kg)	294 – 312	295 – 330
Conductivity (mS/cm)	16,98 – 17,93	17,75 – 18,25
CBC Performance *	Passes test	Passes test
Particles (PLT/WBC channels)	≤ 5	≤ 0,2

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual values meet specifications.

Date: 2024-08-13


Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

