


Certificate of Analysis

Product BOULE EASYCLEANER (50 mL)
Product No 1504513
Lot No. 2024071601
Manufacturing date 2024-07-16
Expiration Date 2026-07-15

Test Hypochlorite	Specification
Hypochlorite concentration	4.50 – <5 %
pH	12.60 – 13.20
Blank	PLT ≤ 10 x10 ⁹ /L RBC ≤ 0.02 x 10 ¹² /L

Actual values meet specifications.

Date: 2024-08-20



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

