

# Certificate of Analysis

**Product** MEDONIC M-SERIES DILUENT, RFID (20 L)  
**Product No** 1504460  
**Lot No** 2408-236  
**Manufacturing Date** 2024-08-26  
**Expiration Date** 2027-08-25

Test	Specification
pH	6,90 – 7,10
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,50 – 18,00
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2024-08-29



*Sofia Karlström*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

