

CERTIFICATE OF ANALYSIS

Product: Swelab™AlfaDiluent

Lot No: 2408-008

Product Part #: 501-208


Exp. Date: 2027-07-31

Manufacturing Date: 08-22-2024

Quality Testing Date: 08-28-2024

Parameter	Specification	Result	Pass/Fail
pH (units):	6.50 – 6.90	6.79	Pass
Osmolality (mOsm/kg):	294 – 312	299	Pass
Conductivity (mS/cm):	16.98 – 17.93	17.21	Pass
Background (x10³/µL):	0.0 – 8.0	2.5	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality Assurance

09/03/24

Date

