

CERTIFICATE OF ANALYSIS

Product: Swelab™ AlfaLyse RFID

Lot No: 2408-005

Product Part #: 501-466

Exp. Date: 2027-07-31

Manufacturing Date: 2024-08-26

Quality Testing Date: 2024-08-28

Parameter	Specification	Result	Pass/Fail
pH (units):	7.54 – 8.30	8.25	Pass
Osmolality (mOsm/kg):	295 – 330	318	Pass
Conductivity (mS/cm):	18.00 – 18.50	18.18	Pass
Functional Test	Pass/Fail	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality Assurance

09/03/24

Date

