

LOT: 12408-14  
EXP: 2024-11-05  
Open Vial Stability 5 days

### EXIGO

	WBC 10 <sup>9</sup> /L	RBC 10 <sup>12</sup> /L	Hgb g/dL	MCV fL	PLT 10 <sup>9</sup> /L
System specific values	8.6	4.12	11.5	77.5	208
Tolerance limits ±	0.3	0.10	0.2	2.0	15

This assay sheet is also available on :  
[www.boule.com](http://www.boule.com)

## For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo  
Lot: 12408-14  
Prod.: Vet-Cal  
ID: 1240814+  
Exp.: 5-Nov-2024

