

## CERTIFICATE OF ANALYSIS

**Product:** Medonic™ M-Series Diluent with RFID

**Lot No:** 2409-003

**Product Part #:** 501-463

**Exp. Date:** 2027-08-31

**Manufacturing Date:** 2024-09-20

**Quality Testing Date:** 2024-09-27

Parameter	Specification	Result	Pass/Fail
pH (units):	6.90 – 7.20	7.14	Pass
Osmolality (mOsm/kg):	290 – 310	292	Pass
Conductivity (mS/cm):	17.50 – 18.00	17.54	Pass
Background Test (x10 <sup>3</sup> /μL):	0.0 – 5.0	0.3	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality and Regulatory



Date

