

# Certificate of Analysis

**Product** MEDONIC M-SERIES DILUENT (20 L)  
**Product No** 1504122  
**Lot No** 2409-245  
**Manufacturing Date** 2024-09-30  
**Expiration Date** 2027-09-29

| Test                        | Specification |
|-----------------------------|---------------|
| pH                          | 6,90 – 7,10   |
| Osmolality (mOsm/kg)        | 290 – 310     |
| Conductivity (mS/cm)        | 17,50 – 18,00 |
| Particles (PLT/RBC channel) | ≤ 5           |

Actual value meets specification.

Date: 2024-10-01



*Ulrika Miefalk*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

